

CERTIFICATE

Choice Calibration Certificate#:12.1921

Firm:FASTEC SERVICES
Addr:469 LINDBURG STREET
City:CHARLESTON
State:SC, 29412

Cal Date:01/31/2012

Choice #: F06-010

Mfg:SALEM BALL Mod:5.5 MM
A/N:
S/N:LOT#90633 Dept:

STEEL BALLS, SS316

Loc :

Condition
Received:In Tolerance
Returned:In Tolerance

Procedure Used
QTY20+/- .005MM
Accuracy:GRADE:100

Environment
Temp.:21°C
Humid:40%

Reference Standards				Accuracy	Traceability	Cal
A/N	Mfg.	Model	Cal.Due Date	Uncertny	Number	Ratio
103	MTI	516-901	01/04/4013	GRADE 1	LD7-0204/172-10	>=4:1
110	SOKKI	225-002	02/01/2012	0.00500%	821/276614-08	>=4:1

Calibration Data: As Found & As Left.
VERIFIED QTY:20, 5.5 MM, SS316 BALLS TO WITHIN +/- .005 MM

Calibration History:

No Histories on File

Choice Calibration Certifies this Instrument meets the stated accuracy with test results relating to this instrument using Manufacturer / Industry Techniques, or Customer Requirements. Services designed to comply with as required: ISO 9001:2008, ISO 10012 ISO 17025:2005 ANSI/NC SL Z540 former MIL-STD-45662A Using Certified Standards Traceable to N.I.S.T. as Required Uncertainties Estimated at a 95% Confidence Level (k=2).

No reproduction except in full without written approval.

Choice Calibration Company 17411 Irvine Blvd. Unit C
Tustin, CA 92780 Ph: 714-730-1670, Fax: 714-730-7818

Q.C.
11

QA Approval, By:Ken Meadows

Issued Date:01/31/2012