

CERTIFICATE

Choice Calibration Certificate#:12.2669

Firm:FASTEC SERVICES
Addr:469 LINDBURG STREET
City:CHARLESTON
State:SC, 29412

Cal Date:03/23/2012

Choice #: F06-016

Mfg:SALEM BALL Mod:2.00 MM
A/N:
S/N:LOT#SS78139 Dept:

TEST SPHERES, SS304

Loc :

| Condition | Procedure Used | Environment |
|-----------------------|--------------------|-------------|
| Received:In Tolerance | QTY20+/- .005MM | Temp.:20°C |
| Returned:In Tolerance | Accuracy:GRADE:100 | Humid:42% |

| Reference Standards | | | | Accuracy | Traceability | Cal |
|---------------------|-------|---------|--------------|-----------|-----------------|-------|
| A/N | Mfg. | Model | Cal.Due Date | Uncertnry | Number | Ratio |
| 103 | MTI | 516-901 | 01/04/4013 | GRADE 1 | LD7-0204/172-10 | >=4:1 |
| 110 | SOKKI | 225-002 | 02/01/2013 | 0.00500% | 681/279723-10 | >=4:1 |

Calibration Data: As Found & As Left.
VERIFIED QTY:20, 2.0 MM TEST SPHERES WITHIN +/- .005 MM

Calibration History:

No Histories on File

Choice Calibration Certifies this Instrument meets the stated accuracy with test results relating to this instrument using Manufacturer / Industry Techniques, or Customer Requirements. Services designed to comply with as required: ISO 9001:2008, ISO 10012 ISO 17025:2005 ANSI/NCSL Z540 former MIL-STD-45662A Using Certified Standards Traceable to N.I.S.T. as Required Uncertainties Estimated at a 95% Confidence Level (k=2).

No reproduction except in full without written approval.

Choice Calibration Company 17411 Irvine Blvd. Unit C
Tustin, CA 92780 Ph: 714-730-1670, Fax: 714-730-7818

Q.C.
11

QA Approval, By:Ken Meadows

Issued Date:03/23/2012