

CERTIFICATE

Choice Calibration Certificate#:12.2970

Firm:FASTEC SERVICES
Addr:469 LINDBURG STREET
City:CHARLESTON
State:SC, 29412

Cal Date:04/12/2012

Choice #: F06-019

Mfg:SALEM BALL

Mod:1.20 MM

TEST SPHERES, SS316

A/N:

S/N: CERT#0786313

Dept:

Loc :

Condition

Received:In Tolerance

Returned:In Tolerance

Procedure Used

QTY:20-.000031

Accuracy:GRADE:100

Environment

Temp.:21°C

Humid:47%

Reference Standards

A/N	Mfg.	Model	Cal.Due Date	Accuracy Uncertny	Traceability Number	Cal Ratio
-----	------	-------	--------------	----------------------	------------------------	--------------

111	HEIDN	VRZ404	02/01/2013	0.0001%	681/279723-10	>=4:1
-----	-------	--------	------------	---------	---------------	-------

103	MTI	516-901	01/04/4013	GRADE 1	LD7-0204/172-10	>=4:1
-----	-----	---------	------------	---------	-----------------	-------

Calibration Data: As Found & As Left.

VERIFIED QTY:20, 1.20 MM 316SS TEST SPHERES WITHIN -.000031

Calibration History:

No Histories on File

Choice Calibration Certifies this Instrument meets the stated accuracy with test results relating to this instrument using Manufacturer / Industry Techniques, or Customer Requirements. Services designed to comply with as required: ISO 9001:2008, ISO 10012 ISO 17025:2005 ANSI/NCSL Z540 former MIL-STD-45662A Using Certified Standards Traceable to N.I.S.T. as Required Uncertainties Estimated at a 95% Confidence Level (k=2).

No reproduction except in full without written approval.

Choice Calibration Company 17411 Irvine Blvd. Unit C
Tustin, CA 92780 Ph: 714-730-1670, Fax: 714-730-7818

Q.C.
11

QA Approval, By:Ken Meadows

Issued Date:04/12/2012