

CERTIFICATE

Choice Calibration Certificate#:12.3510

Firm:FASTEC SERVICES
Addr:3226 MAYBANK HWY, STE:E5
City:JOHNS ISLAND
State:SC, 29455

Cal Date:05/31/2012

Choice #: F06-024

Mfg:SALEM BALL Mod:1.50 MM
A/N:

TEST SPHERE, SS440

S/N:LOT# 101073854 Dept:

Loc :

Condition	Procedure Used	Environment
Received:In Tolerance	QTY20+/- .002MM	Temp.:20°C
Returned:In Tolerance	Accuracy:GRADE:25	Humid:40%

Reference Standards				Accuracy	Traceability	Cal
A/N	Mfg.	Model	Cal.Due Date	Uncertny	Number	Ratio
103	MTI	516-901	01/04/4013	GRADE 1	LD7-0204/172-10	>=4:1
111	HEIDN	VRZ404	02/01/2013	0.0001%	681/279723-10	>=4:1

Calibration Data: As Found & As Left.
VERIFIED QTY:20 1.50MM TEST SPHERES WITHIN +/- .002MM

Calibration History:

No Histories on File

Choice Calibration Certifies this Instrument meets the stated accuracy with test results relating to this instrument using Manufacturer / Industry Techniques, or Customer Requirements. Services designed to comply with as required: ISO 9001:2008, ISO 10012 ISO 17025:2005 ANSI/NCSL Z540 former MIL-STD-45662A Using Certified Standards Traceable to N.I.S.T. as Required Uncertainties Estimated at a 95% Confidence Level (k=2). No reproduction except in full without written approval. Choice Calibration Company 17411 Irvine Blvd. Unit C Tustin, CA 92780 Ph: 714-730-1670, Fax: 714-730-7818



QA Approval, By:Ken Meadows

Issued Date:05/31/2012