

# CERTIFICATE

Choice Calibration Certificate#:12.3512

Firm:FASTEC SERVICES  
Addr:3226 MAYBANK HWY, STE:E5  
City:JOHNS ISLAND  
State:SC, 29455

Cal Date:05/31/2012

Choice #: F06-026

Mfg:SALEM BALL Mod:3.00 MM

TEST SPHERE, SS304

A/N:

S/N:LOT# 2C02-02

Dept:

Loc :

Condition	Procedure Used	Environment
Received:In Tolerance	QTY20+/- .005MM	Temp.:20°C
Returned:In Tolerance	Accuracy:GRADE:100	Humid:40%

Reference Standards				Accuracy	Traceability	Cal
A/N	Mfg.	Model	Cal.Due Date	Uncertny	Number	Ratio
103 MTI	516-901	01/04/4013	GRADE 1	LD7-0204/172-10	>=4:1	
111 HEIDN	VRZ404	02/01/2013	0.0001%	681/279723-10	>=4:1	

Calibration Data: As Found & As Left.  
VERIFIED QTY:20, 3.00MM TEST SPHERES WITHIN +/- .005MM

### Calibration History:

No Histories on File

Choice Calibration Certifies this Instrument meets the stated accuracy with test results relating to this instrument using Manufacturer / Industry Techniques, or Customer Requirements. Services designed to comply with as required: ISO 9001:2008, ISO 10012 ISO 17025:2005 ANSI/NCSL Z540 former MIL-STD-45662A Using Certified Standards Traceable to N.I.S.T. as Required Uncertainties Estimated at a 95% Confidence Level (k=2). No reproduction except in full without written approval. Choice Calibration Company 17411 Irvine Blvd. Unit C Tustin, CA 92780 Ph: 714-730-1670, Fax: 714-730-7818

Q.C.  
11

QA Approval, By:Ken Meadows

Issued Date:05/31/2012