

CERTIFICATE

Choice Calibration Certificate#:12.3513

Firm:FASTEC SERVICES
Addr:3226 MAYBANK HWY, STE:E5
City:JOHNS ISLAND
State:SC, 29455

Cal Date:05/31/2012

Choice #: F06-027

Mfg:SALEM BALL

Mod:4.50 MM

STEEL BALLS, FE

A/N:

S/N:

Dept:

Loc :

Condition
Received:In Tolerance
Returned:In Tolerance

Procedure Used
QTY20+/- .005MM
Accuracy:GRADE:100

Environment
Temp.:20°C
Humid:40%

Reference Standards

A/N	Mfg.	Model	Cal.Due Date	Accuracy Uncertny	Traceability Number	Cal Ratio
103	MTI	516-901	01/04/4013	GRADE 1	LD7-0204/172-10	>=4:1
111	HEIDN	VRZ404	02/01/2013	0.0001%	681/279723-10	>=4:1

Calibration Data: As Found & As Left.

VERIFIED QTY:20, 4.50MM STEEL BALLS WITHIN +/- .005MM

Calibration History:

No Histories on File

Choice Calibration Certifies this Instrument meets the stated accuracy with test results relating to this instrument using Manufacturer / Industry Techniques, or Customer Requirements. Services designed to comply with as required: ISO 9001:2008, ISO 10012 ISO 17025:2005 ANSI/NCSL Z540 former MIL-STD-45662A Using Certified Standards Traceable to N.I.S.T. as Required Uncertainties Estimated at a 95% Confidence Level (k=2).

No reproduction except in full without written approval.

Choice Calibration Company 17411 Irvine Blvd. Unit C
Tustin, CA 92780 Ph: 714-730-1670, Fax: 714-730-7818

Q.C.
11

QA Approval, By:Ken Meadows

Issued Date:05/31/2012