

# CERTIFICATE

Choice Calibration Certificate#:14.1599

Firm:FASTEC SERVICES  
Addr:3226 MAYBANK HWY, STE:E5  
City:JOHNS ISLAND  
State:SC, 29455

Cal Date:01/10/2014

Choice #: F06-044

Mfg:SALEM BALL Mod:15.0 MM  
A/N:  
S/N:LOT# 66144 Dept:

STEEL BALLS, 316 SS

Loc :

Condition	Procedure Used	Environment
Received:In Tolerance	QTY20+/- .010MM	Temp.:20°C
Returned:In Tolerance	Accuracy:GRADE:100	Humid:42%

Reference Standards	Accuracy	Traceability	Cal
A/N   Mfg.   Model   Cal.Due Date	Uncertnry	Number	Ratio
103 MTI 516-901 01/04/2015	GRADE 1	LD7-0204/172-10	>=4:1
110 SOKKI 225-002 02/01/2014	0.00500%	681/279723-10	>=4:1

Calibration Data: As Found & As Left.  
VERIFIED QTY:20, 15.0 MM, 316SS BALLS TO WITHIN +/- .010 MM.

Calibration History:

No Histories on File

Choice Calibration Certifies this Instrument meets the stated accuracy with test results relating to this instrument using Manufacturer / Industry Techniques, or Customer Requirements. Services designed to comply with as required: ISO 9001:2008, ISO 10012 ISO 17025:2005 ANSI/NCSL Z540 former MIL-STD-45662A Using Certified Standards Traceable to N.I.S.T. as Required Uncertainties Estimated at a 95% Confidence Level (k=2).

No reproduction except in full without written approval.

Choice Calibration Company 17411 Irvine Blvd. Unit C  
Tustin, CA 92780 Ph: 714-730-1670, Fax: 714-730-7818

Q.C.  
11

QA Approval, By:Ken Meadows

Issued Date:01/10/2014