

CERTIFICATE

Choice Calibration Certificate#:14.5385

Firm:FASTEC SERVICES
Addr:3226 MAYBANK HWY, STE:E5
City:JOHNS ISLAND
State:SC, 29455

Cal Date:09/13/2014

Choice #: F06-045

Mfg:SALEM BALL Mod:2.00 MM
A/N:
S/N:LOT# 130452 Dept:

STEEL BALLS, SS316

Loc :

Condition	Procedure Used	Environment
Received:In Tolerance	QTY20+/- .005MM	Temp.:20°C
Returned:In Tolerance	Accuracy:GRADE:100	Humid:44%

Reference Standards				Accuracy	Traceability	Cal
A/N	Mfg.	Model	Cal.Due Date	Uncertny	Number	Ratio
103	MTI	516-901	01/04/2015	GRADE 1	LD7-0204/172-10	>=4:1
110	SOKKI	225-002	02/01/2015	0.00500%	681/279723-10	>=4:1

Calibration Data: As Found & As Left.
VERIFIED QTY:20 2.00MM STEEL BALLS WITHIN +/- .005MM

Calibration History:

No Histories on File

Choice Calibration Certifies this Instrument meets the stated accuracy with test results relating to this instrument using Manufacturer / Industry Techniques, or Customer Requirements. Services designed to comply with as required: ISO 9001:2008, ISO 10012 ISO 17025:2005 ANSI/NC SL Z540 former MIL-STD-45662A Using Certified Standards Traceable to N.I.S.T. as Required Uncertainties Estimated at a 95% Confidence Level (k=2).

No reproduction except in full without written approval.

Choice Calibration Company 17411 Irvine Blvd. Unit C
Tustin, CA 92780 Ph: 714-730-1670, Fax: 714-730-7818

Q.C.
11

QA Approval, By:Ken Meadows

Issued Date:09/13/2014