

# CERTIFICATE

Choice Calibration Certificate#:15.1743

Firm:FASTEC SERVICES  
Addr:3226 MAYBANK HWY, STE:E5  
City:JOHNS ISLAND  
State:SC, 29455

Cal Date:02/12/2015

Choice #: F06-050

Mfg:SALEM BALL Mod:4.50 MM  
A/N:  
S/N:LOT#13012001 Dept:

STEEL BALLS, SS304

Loc :

Condition	Procedure Used	Environment
Received:In Tolerance	QTY20+/- .005MM	Temp.:20°C
Returned:In Tolerance	Accuracy:GRADE:100	Humid:40%

Reference Standards				Accuracy	Traceability	Cal
A/N	Mfg.	Model	Cal.Due Date	Uncertny	Number	Ratio
103 MTI	516-901	01/09/2017	GRADE 1	RLD8-0318/14913	>=4:1	
110 SOKKI	225-002	02/01/2016	0.00500%	RLD8-0318/14913	>=4:1	

Calibration Data: As Found & As Left.  
VERIFIED QTY: 20, 4.50 MM SS304 BALLS ARE WITHIN +/- .005MM

## Calibration History:

No Histories on File

Choice Calibration Certifies this Instrument meets the stated accuracy with test results relating to this instrument using Manufacturer / Industry Techniques, or Customer Requirements. Services designed to comply with as required: ISO 9001:2008, ISO 10012 ISO 17025:2005 ANSI/NCSL Z540 former MIL-STD-45662A Using Certified Standards Traceable to N.I.S.T. as Required Uncertainties Estimated at a 95% Confidence Level (k=2). No reproduction except in full without written approval. Choice Calibration Company 17411 Irvine Blvd. Unit C Tustin, CA 92780 Ph: 714-730-1670, Fax: 714-730-7818

Q.C.  
11

QA Approval, By:Ken Meadows

Issued Date:02/12/2015