

CERTIFICATE

Choice Calibration Certificate#:15.2026

Firm:FASTEC SERVICES
Addr:3226 MAYBANK HWY, STE:E5
City:JOHNS ISLAND
State:SC, 29455

Cal Date:02/25/2015

Choice #: F06-051

Mfg:SALEM BALL Mod:4.50 MM
A/N:
S/N:LOT# 72099 Dept:

STEEL BALLS, BRASS

Loc :

Condition	Procedure Used	Environment
Received:In Tolerance	QTY20+/- .013MM	Temp.:20°C
Returned:In Tolerance	Accuracy:GRADE:200	Humid:40%

Reference Standards				Accuracy	Traceability	Cal
A/N	Mfg.	Model	Cal.Due Date	Uncertnly	Number	Ratio
103	MTI	516-901	01/09/2017	GRADE 1	RLD8-0318/14913	>=4:1
110	SOKKI	225-002	02/01/2016	0.00500%	RLD8-0318/14913	>=4:1

Calibration Data: As Found & As Left.
QTY: 20, 4.50MM BRASS BALLS VERIFIED TO WITHIN:+/- .013MM

Calibration History:

No Histories on File

Choice Calibration Certifies this Instrument meets the stated accuracy with test results relating to this instrument using Manufacturer / Industry Techniques, or Customer Requirements. Services designed to comply with as required: ISO 9001:2008, ISO 10012 ISO 17025:2005 ANSI/NC SL Z540 former MIL-STD-45662A Using Certified Standards Traceable to N.I.S.T. as Required
Uncertainties Estimated at a 95% Confidence Level (k=2).
No reproduction except in full without written approval.
Choice Calibration Company 17411 Irvine Blvd. Unit C
Tustin, CA 92780 Ph: 714-730-1670, Fax: 714-730-7818



QA Approval, By:Ken Meadows

Issued Date:02/25/2015