

CERTIFICATE

Choice Calibration Certificate#:15.6267

Firm:FASTEC SERVICES
Addr:132 SLATE STONE DRIVE UNIT A
City:SUMMERVILLE
State:SC, 29486

Cal Date:11/25/2015

Choice #: F06-052

Mfg:SALEM BALL Mod:2.38 MM
A/N:
S/N:LOT# 1790-1 Dept:

STEEL BALLS, SS316

Loc :

| Condition | Procedure Used | Environment |
|-----------------------|--------------------|-------------|
| Received:In Tolerance | QTY20+/- .005MM | Temp.:20°C |
| Returned:In Tolerance | Accuracy:GRADE:200 | Humid:42% |

| Reference Standards | Accuracy | Traceability | Cal |
|-----------------------------------|----------|-----------------|-------|
| A/N Mfg. Model Cal.Due Date | Uncertny | Number | Ratio |
| 103 MTI 516-901 01/09/2017 | GRADE 1 | RLD8-0318/14913 | >=4:1 |
| 110 SOKKI 225-002 02/01/2016 | 0.00500% | RLD8-0318/14913 | >=4:1 |

Calibration Data: As Found & As Left.
VERIFIED QTY:20, 2.38mm SS316 BALLS ARE WITHIN +/- .005 MM

Calibration History:

No Histories on File

Choice Calibration Certifies this Instrument meets the stated accuracy with test results relating to this instrument using Manufacturer / Industry Techniques, or Customer Requirements. Services designed to comply with as required: ISO 9001:2008, ISO 10012 ISO 17025:2005 ANSI/NCSL Z540 former MIL-STD-45662A Using Certified Standards Traceable to N.I.S.T. as Required Uncertainties Estimated at a 95% Confidence Level (k=2).

No reproduction except in full without written approval.

Choice Calibration Company 17411 Irvine Blvd. Unit C
Tustin, CA 92780 Ph: 714-730-1670, Fax: 714-730-7818

Q.C.
11

QA Approval, By:Ken Meadows

Issued Date:11/25/2015