



Fastec Services, LLC

Industrial Metal Detector Specialists

(864) 836-3600 www.fastecservices.com

750 Circle Dr, Travelers Rest, SC 29690

Certificate of Compliance

Date: 3/03/2022

Size: 5.0mm Soda Lime Glass

Diameter Tolerance: +/-0.03mm

Fastec Services, LLC, further that the sphere was measured by meter compared with Mitutoyo Masters which are traceable to the National Institute of Standards and Technology Test Service ID number 685-O-000000362-19 (Special Test of Spherical Diameter Standards) and has been manufactured in accordance with standards ANSI, AFBMA and ISO3290.

Lot# SLG18903

Silica	72%
Soda	15%
Lime	9%
Magnesium	3%
Alumina	1%

We certify that the materials presented here comply with the above specifications and descriptions.